

To protect the privacy of customers or suppliers of Salamander Jewelry.
We erased their address from this document.

CERTIFICATE OF STERILIZATION

PRODUCT : **ORDER NO. 2103590**

CUSTOMER : Salamander Jewelry Co.,Ltd

CERT. NO : CM 087 /15

We here by certify that following products were sterilized by E.O.at
CITY MEDICAL SUPPLY. The result of sterilization by biological indicator are passed

DATE : 22- 5 -2558

| CODE | Iteme | Color | LOT.NO | TOTAL QUANTITY | STERILIZED DATE | SPECIFICATION | | RESULT | | |
|----------------------------|---|-------|----------|-------------------|--------------------|--------------------|----------------------|--------------------|----------------------|------|
| | | | | | | CHEMICAL TESTED | BIOLOGICAL TESTED | CHEMICAL TESTED | BIOLOGICAL TESTED | |
| TE-Z-TMBN-1.2-12.0-3 | STERILE TITANIUM MICRO CURVED BARBELL FOR TRENDLINE | CR | LOT-0062 | 100 | 21 May 15 | Positive | Negative | Positive | Negative | Pass |
| TE-Z-TMBN-1.2-8.0-3 | STERILE TITANIUM MICRO CURVED BARBELL FOR TRENDLINE | | | 100 | | | | | | |
| TE-Z-TMLB-1.2-8.0-3 | STERILE TITANIUM MICRO LABRET FOR TRENDLINE | | | 200 | | | | | | |
| TE-Z-TBNDJ-1.6-12.0-5/8-CR | TERILE TITANIUM DOUBLE JEWELED NAVEL BANANA FOR TRENDLINE | | | 100 | | | | | | |
| TE-Z-TMLB-1.2-10.0-3 | STERILE TITANIUM MICRO LABRET FOR TRENDLINE | | | 100 | | | | | | |
| Z-PMRP-CRE | STERILE DISPOSAL TOOLS | | | 1315 | | | | | | |

To protect the privacy of customers or suppliers of Salamander Jewelry.
We erased their address from this document.