

To protect the privacy of customers or suppliers of Salamander Jewelry.  
We erased their address from this document.

M/C : SHINVA

Document No. : TR-EO-021310316

MODEL : XG 1 SERIES

Date : 31/3/2016

SERIAL No. : 070997

Page : 2 / 4

PRODUCT : ORDER NO. 2142277

**SAMPLE SLIP PARAMETER**

EXHAUST:04H COMIN TOTAL :12H 21MIN  
PRETT :01H 11MIN STEP :07H 30MIN  
ACTUAL TIME

05:57:07 STERILIZATION SUCCESS  
 END 05:56:56 056,40 -004Pa  
 05:56:56 056,40 -004Pa  
 02:00:25 056,10 -015Pa  
 01:52:25 056,00 -020Pa  
 EXHAUST 01:45:55 056,00 005Pa  
 01:42:25 056,00 005Pa  
 01:32:25 055,90 004Pa  
 01:22:25 056,00 005Pa  
 01:12:25 055,90 005Pa  
 01:02:25 056,00 005Pa  
 00:52:25 056,10 005Pa  
 00:42:25 056,00 005Pa  
 00:32:25 056,00 005Pa  
 00:22:25 055,90 004Pa  
 00:12:25 056,00 005Pa  
 00:02:25 056,00 005Pa  
 23:52:25 055,60 005Pa  
 23:42:25 056,00 005Pa  
 23:32:25 057,10 005Pa  
 23:22:25 056,10 005Pa  
 23:12:25 056,00 005Pa  
 23:02:25 056,10 005Pa  
 22:52:25 056,10 005Pa  
 22:42:25 056,00 005Pa  
 22:32:25 055,90 005Pa  
 22:22:25 056,00 005Pa  
 22:12:25 056,00 004Pa  
 22:02:25 056,00 005Pa

21:52:25 056,00 005Pa  
 21:42:25 055,90 004Pa  
 21:32:25 056,00 005Pa  
 21:22:25 056,00 004Pa  
 21:12:25 056,00 005Pa  
 21:02:25 055,90 004Pa  
 20:52:25 056,00 005Pa  
 20:42:25 055,60 005Pa  
 20:32:25 056,00 007Pa  
 20:22:25 056,00 000Pa  
 20:12:25 056,00 004Pa  
 20:02:25 056,00 006Pa  
 19:52:25 056,00 000Pa  
 19:42:25 055,70 006Pa  
 19:32:25 055,70 006Pa  
 19:22:25 055,70 004Pa  
 19:12:25 056,00 005Pa  
 19:02:25 055,20 247Pa  
 18:52:25 055,60 311Pa  
 STERI 18:46:47 055,90 220Pa  
 18:42:25 056,10 -026Pa  
 ADD ED 18:41:11 055,60 -069Pa  
 18:32:25 055,90 -020Pa  
 HUMIDI 18:31:08 055,60 -069Pa  
 PRETT 18:22:25 055,60 -001Pa  
 PREHEAT 17:36:30 051,20 -002Pa  
 STATE TIME TEMP PRE  
 RUN COUNT : 0610  
 ED PRE :050Pa VENT TIME:04H 00M  
 STERI TEMP: 056,00 STERI TIME:07H 30M  
 PRE TYP 17:36:30 051,20 -002Pa  
 START TIME: 2016-03-29 17:36:22