

To protect the privacy of customers or suppliers of Salamander Jewelry.
We erased their address from this document.

M/C : SHINVA

MODEL : XG 1 SERIES

SERIAL No. : 070997

Document No. : TR-EO-028210416

Date : 21/4/2016

Page : 2 / 4

PRODUCT : ORDER NO. 2143741

SAMPLE SLIP PARAMETER

EXHAUST:04H DOMIN TOTAL :12-11MIN
PRETT :00H 59MIN STER :07H 00MIN
ACTUAL TIME

07:31:55 STERILIZATION SUCCESS

END 07:31:44 057,40 -001kPa
07:31:44 057,40 -001kPa
08:36:02 056,90 -015kPa
08:30:14 056,80 -045kPa
08:20:14 056,80 004kPa
EXHAUST 08:19:44 057,00 005kPa
08:10:14 056,80 004kPa
08:00:14 056,80 004kPa
02:50:14 056,90 005kPa
02:40:14 056,90 005kPa
02:30:14 056,90 005kPa
02:20:14 057,10 005kPa
02:10:14 057,10 005kPa
02:00:14 057,10 004kPa
01:50:14 057,10 004kPa
01:40:14 057,10 004kPa
01:30:14 056,90 005kPa
01:20:14 057,10 005kPa
01:10:14 057,10 004kPa
01:00:14 057,10 004kPa
00:50:14 057,10 004kPa
00:40:14 057,00 004kPa
00:30:14 057,10 004kPa
00:20:14 057,10 005kPa
00:10:14 056,90 004kPa
00:00:14 057,00 004kPa
23:50:14 057,00 004kPa
23:40:14 057,10 004kPa

23:30:14 057,10 004kPa
23:20:14 056,70 003kPa
23:10:14 057,10 004kPa
23:00:14 057,00 005kPa
22:50:14 057,00 005kPa
22:40:14 057,20 004kPa
22:30:14 057,10 006kPa
22:20:14 057,30 004kPa
22:10:14 057,10 007kPa
22:00:14 057,20 002kPa
21:50:14 057,10 004kPa
21:40:14 057,00 006kPa
21:30:14 057,00 007kPa
21:20:14 056,40 005kPa
21:10:14 056,90 004kPa
21:00:14 056,80 004kPa
20:50:14 057,00 004kPa
20:40:14 056,90 005kPa
20:30:14 056,90 011kPa
20:20:14 057,10 020kPa
STERI 20:19:56 057,10 012kPa
ADD EO 20:14:14 056,90 -004kPa
20:10:14 057,00 -018kPa
HUMIDI 20:04:10 056,70 -007kPa
20:00:14 056,70 -014kPa
PRETT 19:50:14 056,50 001kPa
PREHEAT 19:20:57 056,90 000kPa
STATE TIME TEMP PPE

RUN COUNT : 0634

EO PRE :060kPa VENT TIME:04-00h

STERI TEMP: 056,00 STERI TIME:07H 00M

PRO TYP 19:20:57 056,90 000kPa

START TIME: 2016-04-19 19:20:49