

To protect the privacy of customers or suppliers of Salamander Jewelry.  
We erased their address from this document.

SAMPLE SLIP PARAMETER

EXHAUST:04H DOWIN TOTAL :12H 10MIN  
PRETT :01H 05MIN STERI :07H 00MIN  
ACTUAL TIME

06:05:30 STEILIZATION SUCCESS

END 06:05:19 059,50 -005kPa  
06:05:19 059,50 -005kPa  
02:08:47 058,80 -015kPa  
02:04:49 059,10 -016kPa  
EXHAUST 01:59:19 060,00 -001kPa  
01:54:49 059,20 -001kPa  
01:44:49 059,20 -001kPa  
01:34:49 059,20 -001kPa  
01:24:49 059,20 -001kPa  
01:14:49 059,20 -001kPa  
01:04:49 059,30 -001kPa  
00:54:49 059,30 -001kPa  
00:44:49 059,20 -002kPa  
00:34:49 059,30 001kPa  
00:24:49 059,30 -001kPa  
00:14:49 059,70 -001kPa  
00:04:49 059,20 -001kPa  
23:54:49 059,10 -001kPa  
23:44:49 059,10 -001kPa  
23:34:49 059,90 -001kPa  
23:24:49 059,20 -001kPa  
23:14:49 059,20 -001kPa  
23:04:49 059,20 -001kPa  
22:54:49 059,20 000kPa  
22:44:49 059,00 -001kPa  
22:34:49 059,20 000kPa  
22:24:49 059,10 000kPa  
22:14:49 059,20 001kPa

22:04:49 059,20 -001kPa  
21:54:49 059,00 000kPa  
21:44:49 059,00 000kPa  
21:34:49 059,10 -001kPa  
21:24:49 059,00 000kPa  
21:14:49 058,90 000kPa  
21:04:49 058,90 000kPa  
20:54:49 058,90 000kPa  
20:44:49 058,80 -001kPa  
20:34:49 058,70 000kPa  
20:24:49 058,60 001kPa  
20:14:49 058,50 000kPa  
20:04:49 058,30 000kPa  
19:54:49 058,20 054kPa  
19:44:49 058,00 062kPa  
19:34:49 058,10 062kPa  
19:24:49 058,40 140kPa  
19:14:49 058,50 280kPa  
19:04:49 058,80 311kPa  
STERI 19:00:16 058,70 275kPa  
18:54:49 058,90 -029kPa  
ADD ED 18:51:09 058,80 -078kPa  
18:44:49 058,50 -010kPa  
HUMID 18:41:06 058,40 -072kPa  
PRETT 18:34:49 058,10 000kPa  
PREHEAT 17:55:09 053,80 -001kPa  
STATE TIME TEMP PRE

=====  
RUN COUNT : 0792  
ED PRE :060kPa VENT TIME:04H 00M  
STERI TEMP: 056,00 STERI TIME:07H 00M  
PRO TYP 17:55:09 053,80 -001kPa  
START TIME: 2016-08-09 17:55:02