

To protect the privacy of
customers or suppliers of
Salamander Jewelry.
We erased their address from
this document.

Document No. : TR-EO-009270816

Date : 27/8/2016

Page : 2 / 5

PRODUCT : ORDER NO. 2157560

SAMPLE SLIP PARAMETER

EXHAUST:04H 00MIN TOTAL :11H 54MIN

PRETT :00H 52MIN STERI :07H 00MIN

ACTUAL TIME

05:17:42 STEILIZATION SUCCESS

END 05:17:31 059,70 000kPa

05:17:31 059,70 000kPa

01:19:21 059,30 -015kPa

EXHAUST 01:14:31 059,60 -001kPa

01:13:01 059,50 000kPa

01:03:01 059,40 000kPa

00:53:01 059,60 001kPa

00:43:01 059,60 000kPa

00:33:01 059,60 000kPa

00:23:01 059,60 -001kPa

00:13:01 059,40 000kPa

00:03:01 059,30 -001kPa

23:53:01 059,40 -001kPa

23:43:01 059,50 000kPa

23:33:01 059,30 -001kPa

23:23:01 059,50 000kPa

23:13:01 059,20 000kPa

23:03:01 059,40 000kPa

22:53:01 059,50 000kPa

22:43:01 059,30 000kPa

22:33:01 059,30 000kPa

22:23:01 059,30 000kPa

22:13:01 059,20 000kPa

22:03:01 059,00 000kPa

21:53:01 059,10 -001kPa

21:43:01 059,10 000kPa

21:33:01 059,10 000kPa

21:23:01 059,20 000kPa

21:13:01 059,20 -002kPa

21:03:01 059,40 -001kPa

20:53:01 059,40 -001kPa

20:43:01 059,70 000kPa

20:33:01 060,00 -001kPa

20:23:01 060,10 000kPa

20:13:01 060,40 000kPa

20:03:01 060,70 -001kPa

19:53:01 060,70 000kPa

19:43:01 060,80 000kPa

19:33:01 060,80 -001kPa

19:23:01 060,80 000kPa

19:13:01 060,90 052kPa

19:03:01 060,60 061kPa

18:53:01 060,40 076kPa

18:43:01 060,30 123kPa

18:33:01 060,20 311kPa

18:23:01 060,50 311kPa

STERI 18:15:10 060,80 311kPa

18:13:01 060,80 002kPa

ADD EO 18:09:06 060,60 -088kPa

18:03:01 060,50 -005kPa

HUMID 17:59:03 060,30 -088kPa

PRETT 17:53:01 059,40 000kPa

PREHEAT 17:23:54 057,90 -002kPa

STATE TIME TEMP PRE

=====

RUN COUNT : 0817

EO PRE :060kPa VENT TIME:04H 00M

STERI TEMP: 066,0; STERI TIME:07H 00M

PRO TYP 17:23:54 057,90 -002kPa

START TIME: 2016-08-25 17:23:47