

To protect the privacy of customers or suppliers of Salamander Jewelry.
We erased their address from this document.

TEST REPORT

CUSTOMER : Salamander Jewelry Co.,Ltd.

M/C : SHINVA

Document No. : TR-EO-001120121

MODEL : XG 1 SERIES

Date : 12/1/2021

SERIAL No. : 070997

Page : 1 / 4

PRODUCT : ORDER NO. 2331251

Determination of Sterilization Specification

Item Detail	Specification	Setting	Result
Chamber Setting Temperature	55.0 - 59.0 (°C)	59 (°C)	-
Vacuum Level	5-12 kPa	9 kPa	-
Vacuum Time (min)	50-60 min	60 min	-
Humidification			
Humidity (%)	50%-90%	87%	-
Sterilization			
Steri Temp	55.0 - 59.0 (°C)	55.0 (°C)	-
Jacket Temp	60.0 - 65.0 (°C)	65.0 (°C)	-
Gas Injection			
Final pressure (kpa(abs))	0.21-0.23 kPa	0.21 kPa	-
Final temperature of vaporizer, Temp 10 (°C)	55.0 - 59.0 (°C)	55 (°C)	-
Exposure			
Temperature at strat point of Exposure time period Temp. 1 (°C)	55.0 - 59.0 (°C)	60 (°C)	-
Temperature at finish point of Exposure time period Temp. 1 (°C)	55.0 - 59.0 (°C)	55 (°C)	-
Exposure Time (min)	390-420 min	420 min	-
Exhaust			
Exhaust Time (hour)	≥ 4 hour	4 hour	-
First Vacuum Level (kPa)	5-12 kPa	9 kPa	-
Second Vacuum Level (kPa)	5-12 kPa	9 kPa	-
Third Vacuum Level (kPa)	5-12 kPa	9 kPa	-
Storage			
Storage Time	≥ 48 hour	≥ 48 hour	-
Chemical tested, Sterile Strip			
EO sterilization temperatures 55-60 C° Time sterile by EO 7 hours .	Brown	Brown	Pass
Biological Test			
EZ Test Biological Indicator	Brown	Brown	Pass
24 hour results for EO	Red	Red	Pass

To protect the privacy of customers or suppliers of Salamander Jewelry.
We erased their address from this document.

M/C : SHINVA

Document No. : TR-EO-001120121

MODEL : XG 1 SERIES

Date : 12/1/2021

SERIAL No. : 070997

Page : 2 / 4

PRODUCT : ORDER NO. 2331251

SAMPLE SLIP PARAMETER

EXHAUST:04H 00MIN TOTAL :11H 50MIN			
PRETT :00H 48MIN STERI :07H 00MIN			
ACTUAL TIME			
05:09:31 STEILIZATION SUCCESS			
END	05:09:20	061.10	002kPa
	05:09:20	061.10	002kPa
	01:09:27	062.20	-015kPa
	01:07:50	062.30	006kPa
EXHAUST	01:07:20	062.30	007kPa
	09:57:30	062.40	007kPa
	00:47:50	062.40	007kPa
	00:37:50	062.60	007kPa
	00:27:50	062.70	007kPa
	00:17:50	062.70	007kPa
	00:07:50	062.90	007kPa
	23:57:50	062.90	007kPa
	23:47:49	063.10	007kPa
	23:37:49	063.10	007kPa
	23:27:49	063.20	007kPa
	23:17:49	063.20	007kPa
	23:07:49	063.50	007kPa
	22:57:49	063.50	007kPa
	22:47:49	063.60	007kPa
	22:37:49	063.70	007kPa
	22:27:49	063.80	007kPa
	22:17:49	063.80	007kPa
	22:07:49	063.90	007kPa
	21:57:49	064.00	007kPa
	21:47:49	064.10	007kPa
	21:37:49	064.10	007kPa
	21:27:49	064.10	007kPa
	21:17:49	064.20	007kPa

	21:07:49	064.30	007kPa
	20:57:49	064.30	007kPa
	20:47:49	064.30	006kPa
	20:37:49	064.40	006kPa
	20:27:49	064.50	006kPa
	20:17:49	064.50	006kPa
	20:07:49	064.50	006kPa
	19:57:49	064.50	014kPa
	19:47:49	064.50	023kPa
	19:37:49	064.70	026kPa
	19:27:49	064.70	029kPa
	19:17:49	064.80	032kPa
	19:07:49	064.80	043kPa
	18:57:49	064.90	052kPa
	18:47:49	064.90	057kPa
	18:37:49	064.90	089kPa
	18:27:49	065.00	219kPa
	18:17:49	064.90	311kPa
	18:07:49	065.00	311kPa
STERI	18:07:49	065.00	311kPa
ADD ED	18:01:22	065.10	-089kPa
	17:57:49	065.00	001kPa
HUMIDI	17:51:20	064.60	-089kPa
PRETT	17:47:49	064.10	002kPa
PREHEAT	17:19:15	060.00	002kPa
STATE	TIME	TEMP	PRE
RUN COUNT : 0325			
ED PRE :060kPa VENT TIME:04H 00M			
STERI TEMP: 065.00 STERI TIME:07H 00M			
PRE TIP 17:19:15 060.00 002kPa			
START TIME: 2021-01-10 17:19:08			

To protect the privacy of customers or suppliers of Salamander Jewelry. We erased their address from this document.

M/C : SHINVA

Date : 12/1/2021

MODEL : XG 1 SERIES

Page : 3 / 4

SERIAL No. : 070997

PRODUCT : ORDER NO. 2331251

PROCESS FLOW CHART OF STERILIZATION

- > loading > into the chamber > door close > program starts
- > preprocessing > EO entry > sterilization
- > later processing > ventilation > air entry > the end
- > door open > unloading.

We here by certificate that following products were sterilized by E.O. at CITY MEDICAL SUPPLY CO.,LTD .

The result of sterilization by biological indicator are passed.

CODE	Iteme	Color	Lot. No	TOTAL QUANTITY	STERILIZED DATE
TDI-Z-BKMBN-1.2-8.0-3	STERILE BLACK PVD COATED SURGICAL STEEL CURVED MICRO BARBELL FOR TDI			50	
TDI-Z-BL-1.6-12.0-4	STERILE SURGICAL STEEL BARBELL FOR TDI			50	
TDI-Z-BL-1.6-38.0-5	STERILE SURGICAL STEEL BARBELL FOR TDI			50	
TDI-Z-BN-1.6-10.0-4	STERILE SURGICAL STEEL CURVED BARBELL FOR TDI			100	
TDI-Z-BNJ-OP-1.6-10.0-5/8-CR	STERILE SURGICAL STEEL CZECH CRYSTAL JEWELLED NAVEL BANANA FOR TDI			50	
TDI-Z-CB-1.6-10.0-4	STERILE SURGICAL STEEL CIRCULAR BARBELL FOR TDI			100	
TDI-Z-LB-1.6-8.0-4	STERILE SURGICAL STEEL LABRET FOR TDI			100	
TDI-Z-MBL-1.2-10.0-3	STERILE SURGICAL STEEL MICRO BARBELL FOR TDI			50	
TDI-Z-MBL-1.2-9.0-3	STERILE SURGICAL STEEL MICRO BARBELL FOR TDI			50	
TDI-Z-MCB-1.2-12.0-3	STERILE SURGICAL STEEL MICRO CIRCULAR BARBELL FOR TDI			50	
TDI-Z-MCB-1.2-7.0-3	STERILE SURGICAL STEEL MICRO CIRCULAR BARBELL FOR TDI			50	
TDI-Z-MCB-1.2-8.0-3	STERILE SURGICAL STEEL MICRO CIRCULAR BARBELL FOR TDI			200	
TDI-Z-MLB-1.2-10.0-2.5	STERILIZED SURGICAL STEEL MICRO LABRET FOR TDI			100	
TDI-Z-MLBJ-OP-1.2-10.0-3-CR	STERILE SURGICAL STEEL CZECH CRYSTAL JEWELLED MICRO LABRET FOR TDI			100	
TDI-Z-TBL-1.6-12.0-4	STERILE TITANIUM BARBELL FOR TDI			100	
TDI-Z-TBL-1.6-16.0-5	STERILE TITANIUM BARBELL FOR TDI			300	
TDI-Z-TBL-1.6-32.0-5	STERILE TITANIUM BARBELL FOR TDI		LOT-0215	100	10 Jan 21
TDI-Z-TBN-1.6-12.0-4	STERILIZED TITANIUM CURVED BARBELL FOR TDI			50	
TDI-Z-TBNDJ-OP-1.6-10.0-5/8-BZ	STERILE TITANIUM JEWELLED NAVEL BANANA FOR TDI			25	
TDI-Z-TBNJ-OP-1.6-10.0-5/8-CR	STERILE TITANIUM JEWELLED NAVEL BANANA FOR TDI			100	
TDI-Z-TBNJ-OP-1.6-10.0-5/8-RO	STERILE TITANIUM JEWELLED NAVEL BANANA FOR TDI			25	
TDI-Z-TBNN-1.6-10.0-5/8	STERILE TITANIUM NAVEL BANANA FOR TDI			50	
TDI-Z-TINBL-1.6-16.0-5	STERILE TITANIUM INTERNALLY THREADED BARBELL FOR TDI			100	
TDI-Z-TINMLB-1.2-8.0-2	STERILE TITANIUM INTERNALLY THREADED MICRO LABRET FOR TDI			100	
TDI-Z-TMBN-1.2-10.0-3	STERILE TITANIUM MICRO CURVED BARBELL FOR TDI			500	
TDI-Z-TMLB-1.2-12.0-2.5	STERILE TITANIUM MICRO LABRET FOR TDI			200	
TDI-Z-TMLB-1.2-8.0-2.5	STERILE TITANIUM MICRO LABRET FOR TDI			300	
TDI-Z-TMLB-1.2-9.0-2.5	STERILE TITANIUM MICRO LABRET FOR TDI			200	
TDI-Z-TMLBJ-1.2-8.0-2.5-CR	STERILE TITANIUM MICRO LABRET WITH JEWELLED BALL FOR TDI			200	
TDI-Z-TMLBJ-OP-1.2-8.0-3-CR	STERILE TITANIUM MICRO LABRET WITH JEWELLED BALL FOR TDI			400	
TDI-Z-TMLB-1.2-8.0-3	STERILE TITANIUM MICRO LABRET FOR TDI			500	
PT-Z-TBCH-1.2-10.0-4	STERILE TITANIUM BALL CLOSURE RING WITH HEMATITE BALL FOR PTIT ROC			163	
PT-Z-TBCH-1.6-10.0-4	STERILE TITANIUM BALL CLOSURE RING WITH HEMATITE BALL FOR PTIT ROC			108	
PT-Z-TBL-1.6-16.0-5	STERILIZED TITANIUM BARBELL FOR PTIT ROC			83	
PT-Z-TBL-1.6-22.0-5	STERILIZED TITANIUM BARBELL FOR PTIT ROC			161	
PT-Z-TBNDJ-1.6-10.0-5/8-CR	STERILE TITANIUM DOUBLE JEWELLED NAVEL BANANA FOR PTIT ROC			72	

M/C : SHINVA

Document No. : TR-EO-001120121

MODEL : XG 1 SERIES

Date : 12/1/2021

SERIAL No. : 070997

Page : 4 / 4

PRODUCT : ORDER NO. 2331251

CODE	Iteme	Color	Lot. No	TOTAL QUANTITY	STERILIZED DATE
PT-Z-TBNJ-1.6-10.0-5/8-CR	STERILE TITANIUM JEWELLED NAVEL BANANA FOR PTIT ROC			94	
PT-Z-TMBL-1.2-8.0-3-HP	STERILE TITANIUM MICRO BARBELL FOR PTIT ROC			360	
PT-Z-TMBN-1.2-12.0-3	STERILE TITANIUM MICRO CURVED BARBELL FOR PTIT ROC			143	
PT-Z-TMBN-1.2-8.0-3	STERILE TITANIUM MICRO CURVED BARBELL FOR PTIT ROC			126	
PT-Z-TMCB-1.2-10.0-3	STERILE TITANIUM MICRO CIRCULAR BARBELL FOR PTIT ROC			160	
PT-Z-TMCB-1.2-8.0-3	STERILE TITANIUM MICRO CIRCULAR BARBELL FOR PTIT ROC			373	
TDI-Z-TBC-1.2-8.0-4	STERILE TITANIUM BALL CLOSURE RING FOR TDI			100	
TDI-Z-TBC-1.6-12.0-5	STERILE TITANIUM BALL CLOSURE RING FOR TDI			25	
TDI-Z-TBC-1.6-14.0-5	STERILE TITANIUM BALL CLOSURE RING FOR TDI			50	
TDI-Z-TBCH-1.6-12.0-5	STERILE TITANIUM BALL CLOSURE RING WITH HEMATITE BALL FOR TDI			40	
TDI-Z-TCB-1.6-10.0-4	STERILE TITANIUM 90 DEGREE STAPLE BARBELL FOR TDI			100	
TDI-Z-TCB-1.6-8.0-4	STERILIZED TITANIUM CIRCULAR BARBELL FOR TDI			200	
TDI-Z-TMCB-1.2-10.0-3	STERILE TITANIUM MICRO CIRCULAR BARBELL FOR TDI			200	
TDI-Z-TMCB-1.2-8.0-3	STERILE TITANIUM MICRO CIRCULAR BARBELL FOR TDI			100	
THD-Z-BKBL-1.6-22.0-5	STERILE BLACK PVD COATED SURGICAL STEEL BARBELL FOR THENDRO			100	
THD-Z-BKCB-1.6-8.0-4	STERILE BLACK PVD COATED SURGICAL STEEL CIRCULAR BARBELL FOR THENDRO			100	
THD-Z-BKMBL-1.2-6.0-3	STERILE BLACK PVD COATED SURGICAL STEEL MICRO BARBELL FOR THENDRO		LOT-0215	200	10 Jan 21
THD-Z-BKBC-1.2-10.0-4	STERILE BLACK PVD COATED SURGICAL STEEL BALL CLOSURE RING FOR THENDRO			500	
THD-Z-BKBL-1.6-10.0-4	STERILE BLACK PVD COATED SURGICAL STEEL BARBELL FOR THENDRO			300	
THD-Z-BKBL-1.6-14.0-5	STERILE BLACK PVD COATED SURGICAL STEEL BARBELL FOR THENDRO			300	
THD-Z-BKBL-1.6-16.0-5	STERILE BLACK PVD COATED SURGICAL STEEL BARBELL FOR THENDRO			200	
THD-Z-BKCB-1.6-10.0-4	STERILE BLACK PVD COATED SURGICAL STEEL CIRCULAR BARBELL FOR THENDRO			200	
THD-Z-BKCNO-0.8-6.5-2.35	STERILE BLACK PVD COATED SURGICAL STEEL CURVED BALL NOSE STUD FOR THENDRO			200	
THD-Z-BKMBL-1.2-10.0-3	STERILE BLACK PVD COATED SURGICAL STEEL MICRO BARBELL FOR THENDRO			500	
THD-Z-BKMBL-1.2-8.0-3	STERILE BLACK PVD COATED SURGICAL STEEL MICRO BARBELL FOR THENDRO			500	
THD-Z-BKMCB-1.2-10.0-3	STERILE BLACK PVD COATED SURGICAL STEEL MICRO CIRCULAR BARBELLFOR THENDRO			400	
THD-Z-BKMLB-1.2-8.0-3	STERILE BLACK PVD COATED SURGICAL STEEL MICRO LABRETFOR THENDRO			500	
PT-Z-TMLB-1.2-9.0-3	STERILE TITANIUM MICRO LABRET FOR PTIT ROC			486	
PT-Z-TMLB-1.2-9.0-3	STERILE TITANIUM MICRO LABRET FOR PTIT ROC			156	
PT-Z-TMLB-1.2-9.0-3	STERILE TITANIUM MICRO LABRET FOR PTIT ROC			100	
PT-Z-TBCH-1.2-10.0-4	STERILE TITANIUM BALL CLOSURE RING WITH HEMATITE BALL FOR PTIT ROC			115	
PT-Z-TMLB-1.2-9.0-3	STERILE TITANIUM MICRO LABRET FOR PTIT ROC			180	
Z-TJNO-1.0-6.5-2.35-HP-CR	STERILE TITANIUM JEWELLED CURVED NOSE STUD			100	
TDI-Z-LB-1.6-10.0-	STERILE SURGICAL STEEL LABRET FOR TDI			100	
TDI-Z-TMBL-1.2-8.0-3-HP	STERILE TITANIUM MICRO BARBELL FOR TDI			100	
Z-JNO-1.0-6.5-2.35-CR	STERILE SURGICAL STEEL JEWELLED CURVED NOSE STUD			200	
PT-Z-TMLB-1.2-9.0-3	STERILE TITANIUM MICRO LABRET FOR PTIT ROC			292	
PT-Z-TMLBJ-1.2-8.0-2.5-CR	STERILE TITANIUM MICRO LABRET WITH JEWELLED BALL FOR PTIT ROC			60	
Z-TMCB-1.2-8.0-3-HP	STERILE TITANIUM MICRO CIRCULAR BARBELL			50	
Z-GSNEEDLES-0.8-48.0	STERILIZED STAINLESS STEEL NEEDLE			1560	
Z-PRTA-8.0-70.0	STERILE DISPOSABLE MULTI ANGLED RECEIVING TUBE 8.0MM			2000	

To protect the privacy of customers or suppliers of Salamander Jewelry. We erased their address from this document.